

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **TraceMedia™ TCBS Agar/ BacterPlate™ TCBS Agar**
 Product Code: **TM2022.010/ 05022**
 Batch Number: **250924**
 Expiration Date: **251224**
 Storage Conditions: **2°C – 8°C**
 Method: **Quality testing procedure: QT/KT/03.01**
 Execution Date: **25/09/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Smooth, uniform surface, no roughness, no layering	Passed
Color	Light green agar, transparent	Passed
pH	8,6 ± 0,2	Passed
Agar Thickness (mm)	4 ± 1	Passed
Dish Diameter (mm)	90 ± 0,5	Passed
Volume (mL)	20 ± 2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C, for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 18 – 24 hrs, incubated at 35 – 37°C, inoculum ≤ 10³ – 10⁴ CFU/mL		
<i>V. parahaemolyticus</i> ATCC 17802	Good growth, green colonies	Passed
<i>E. coli</i> ATCC 35218	Inhibited, no growth	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

 (Please download the certificate of analysis at the link:: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, September 27, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm




Đoàn Đức Đại