

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **BacterPlate™ Mueller Hinton Agar, Glucose-Methylene Blue**
Product Code: **05033**
Batch Number: **211224**
Expiration Date: **210325**
Storage Conditions: **2°C – 8°C**
Method: **Quality testing procedure: QT/KT/03.01**
Execution Date: **21/12/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Smooth, uniform surface, no roughness, no layering	Passed
Color	Light amber agar, transparent	Passed
pH	7,3 ± 0,2	Passed
Agar Thickness (mm)	4 ± 1	Passed
Dish Diameter (mm)	90 ± 0,5	Passed
Volume (mL)	20 ± 2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 24 – 48 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 24 – 48 hours, incubated at 35 – 37°C, inoculum ≤ 10³ – 10⁴ CFU/mL		
<i>C. albicans</i> ATCC 10231	Good growth	Passed
<i>S. cerevisiae</i> ATCC 20827	Good growth	Passed


CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link:: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm



Đoàn Đức Đại