

## CERTIFICATE OF ANALYSIS

### PRODUCT INFORMATION

Product: **BacterTest™ KOH**  
Product Code: **06004**  
Batch Number: **061224**  
Expiration Date: **061225**  
Storage Conditions: **8°C – 25°C**  
Method: **Quality testing procedure: QT/KT/03.01**  
Execution Date: **06/12/2024**

### RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform, no residue, no impurities, and no stratification	Passed
Color	Clear	Passed
Volume (mL)	100 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
<b>Culture conditions: 18 – 24 hours, incubated at 35 – 37°C, inoculum ≤ 10<sup>2</sup> CFU/mL</b>		
<i>K. pneumonia</i> NCTC 13442	Good growth, Red	Passed
<i>E. coli</i> ATCC 35218	Good growth, Yellow or no change	Passed

### CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 08, 2024

Quality Inspection Personnel

Quality Management Department



Huynh Bich Tram



Doan Duc Dai