

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **TraceMedia™ Stuart Amies/ BacterTube™ Stuart Amies**
Product Code: **TM5922.050/ 08002**
Batch Number: **211224**
Expiration Date: **210325**
Storage Conditions: **2°C – 8°C**
Method: **Quality testing procedure: QT/KT/03.01**
Execution Date: **21/12/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Smooth, uniform surface, no roughness, no layering	Passed
Color	White agar, opaque	Passed
pH	7,2 ± 0,2	Passed
Tube Height (mm)	100 ± 2	Passed
Tube Diameter (mm)	16 ± 0,5	Passed
Volume (mL)	5 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35–37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 18 – 24 hours, incubated at 35–37°C, inoculum ≤ 10² CFU/mL		
<i>S. aureus</i> ATCC 25923	Good growth	Passed
<i>E. coli</i> ATCC 25922	Good growth	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link:: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm



Đoàn Đức Đại