

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **TraceMedia™ Selectine Cysteine/ BacterTube™ Selectine Cysteine**

Product Code: **TM4822.050/ 08005**

Batch Number: **211224**

Expiration Date: **210625**

Storage Conditions: **2°C – 8°C**

Method: **Quality testing procedure: QT/KT/03.01**

Execution Date: **21/12/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform, clear, no sediment, no roughness, and no stratification	Passed
Color	Amber agar, transparent	Passed
pH	7,0 ± 0,2	Passed
Tube Height (mm)	100 ± 2	Passed
Tube Diameter (mm)	16 ± 0,5	Passed
Volume (mL)	5 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 18 – 24 hours, incubated at 35–37°C, 5 – 10% CO2 inoculum ≤ 10² – 10³ CFU/mL		
<i>Salmonella typhimurium</i> ATCC 14028	Good growth	Passed
<i>E.coli</i> ATCC 25922	Inhibited, no growth	Passed


CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link:: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm



Đoàn Đức Đại

LABone Scientific Equipment Co., Ltd.

228/13/3 Nguyen Thi Lang, Tan Phu Trung Commune, Cu Chi District, Ho Chi Minh City, Vietnam

Phone: (+84) 028.6260 2798 Website: <http://www.labone.vn>