

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **TraceMedia™ BHI 20% Glycerol/ BacterTube™ BHI 20% Glycerol**

Product Code: **TM5322.050/ 08012**

Batch Number: **211224**

Expiration Date: **211225**

Storage Conditions: **2°C – 8°C**

Method: **Quality testing procedure: QT/KT/03.01**

Execution Date: **21/12/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Smooth, uniform surface, no roughness, no layering	Passed
Color	Yellow, transparent	Passed
pH	$6,8 \pm 0,2$	Passed
Volume (mL)	$1 \pm 0,2$	Passed
Sterility	No bacterial growth after incubation at 35–37°C for 48–72 hrs	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link:: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm



Đoàn Đức Đại