

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **BacterTube™ Thioglycollate Broth**
 Product Code: **08017**
 Batch Number: **211224**
 Expiration Date: **210625**
 Storage Conditions: **2°C – 8°C**
 Method: **Quality testing procedure: QT/KT/03.01**
 Execution Date: **21/12/24**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform surface, not cloudy, and not layered.	Passed
Color	Light yellow liquid, transparent	Passed
pH	7,1 ± 0,2	Passed
Tube Height (mm)	100 ± 2	Passed
Tube Diameter (mm)	16 ± 0,5	Passed
Volume (mL)	5 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 18 – 24 hours, incubated at 35–37°C, inoculum ≤ 10² CFU/mL		
<i>Clostridium perfringens</i> ATCC 13124	Good growth	Passed
<i>Staphylococcus aureus</i> ATCC 25923	Good growth	Passed
<i>Pseudomonas aeruginosa</i> ATCC 27853	Good growth	Passed
<i>Escherichia coli</i> ATCC 25922	Good growth	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm



Đoàn Đức Đại