

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **BacterTube™ PBS Buffer 1X**
Product Code: **08025**
Batch Number: **211224**
Expiration Date: **210625**
Storage Conditions: **2°C – 8°C**
Method: **Quality testing procedure: QT/KT/03.01**
Execution Date: **21/12/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform surface, not cloudy, and not layered.	Passed
Color	Colorless liquid, transparent	Passed
pH	7,4 ± 0,2	Passed
Tube Height (mm)	100 ± 2	Passed
Tube Diameter (mm)	16 ± 0,5	Passed
Volume (mL)	5 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 24 – 48 hrs	Passed


CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm



Doãn Đức Đại