

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **TraceMedia™ Motility Indole Urease (MIU)/ BacterTube™ Motility Indole Urease (MIU)**
 Product Code: **TM3922.050/ 08029**
 Batch Number: **161224**
 Expiration Date: **160425**
 Storage Conditions: **2°C – 8°C**
 Method: **Quality testing procedure: QT/KT/03.01**
 Execution Date: **16/12/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform surface, non-cloudy, free of impurities, no layering	Passed
Color	Amber solution, transparent	Passed
pH	6,9 ± 0,2	Passed
Tube height (mm)	75 ± 2	Passed
Tube diameter (mm)	12 ± 0,5	Passed
Volume (mL)	2,5 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C, for 18 – 24 hrs	Passed

PERFORMANCE	GROWTH CAPABILITY	MOBILITY	INDOLE	UREASE	RESULTS
Culture conditions: 18 – 24 hrs, incubated at 35 – 37°C					
<i>E.coli</i> ATCC 25922	Good growth	Positive: Stable growth, causing turbidity	Positive: Red ring on the surface of the medium	Negative	Passed
<i>Shigella flexneri</i> ATCC 12022	Good growth	Negative: Stable growth, no turbidity	Negative	Negative	Passed


CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

 (Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 18, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm




Đoàn Đức Đại

LABone Scientific Equipment Co., Ltd.

228/13/3 Nguyen Thi Lang, Tan Phu Trung Commune, Cu Chi District, Ho Chi Minh City, Vietnam

 Phone: (+84) 028.6260 2798 Website: <http://www.labone.vn>

LABone