

BTL/BM/HD/QTTN/01.03

# CERTIFICATE OF ANALYSIS

## PRODUCT INFORMATION

Product: **TraceMedia<sup>TM</sup> Motility Test Medium/ BacterTube<sup>TM</sup> Motility Test Medium**  
 Product Code: **TM4622.050/ 08038**  
 Batch Number: **211224**  
 Expiration Date: **210625**  
 Storage Conditions: **2°C – 8°C**  
 Method: **Quality testing procedure: QT/KT/03.01**  
 Execution Date: **21/12/2024**

## RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform, clear, no sediment, no roughness, and no stratification	Passed
Color	Light amber semi - solid, opaque	Passed
pH	7,0 ± 0,2	Passed
Tube Height (mm)	100 ± 2	Passed
Tube Diameter (mm)	16 ± 0,5	Passed
Volume (mL)	2 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
<b>Culture conditions: 18 – 24 hours, incubated at 35–37°C, inoculum ≤ 10<sup>3</sup> – 10<sup>4</sup> CFU/mL</b>		
<i>Escherichia coli</i> ATCC 25922	Good growth: Bacteria spread out from the stab line.	Passed
<i>Klebsiella pneumonia</i> ATCC 13883	Good growth: Bacteria remain confined to the stab line	Passed

## CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

 (Please download the certificate of analysis at the link:: <https://labone.com.vn/certificate-of-analysis-coa/> )

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm




Đoàn Đức Đại

**LABone Scientific Equipment Co., Ltd.**

228/13/3 Nguyen Thi Lang, Tan Phu Trung Commune, Cu Chi District, Ho Chi Minh City, Vietnam

 Phone: (+84) 028.6260 2798 Website: <http://www.labone.vn>