

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **BacterTube™ MKTTn**
 Product Code: **08049**
 Batch Number: **211224**
 Expiration Date: **210625**
 Storage Conditions: **2°C – 8°C**
 Method: **Quality testing procedure: QT/KT/03.01**
 Execution Date: **21/12/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform, clear, no sediment, no roughness, and no stratification	Passed
Color	Light green coloured opalescent solution	Passed
pH	8,0 ± 0,2	Passed
Tube Height (mm)	100 ± 2	Passed
Tube Diameter (mm)	16 ± 0,5	Passed
Volume (mL)	5 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 18 – 24 hours, incubated at 35 – 37°C, inoculum ≤ 10² CFU/mL, aerobic		
<i>Salmonella Enteritidis</i> WDCM 00030	> 10 colonies	Passed
<i>Salmonella Typhimurium</i> WDCM 00031	> 10 colonies	Passed
<i>Escherichia coli</i> WDCM 00013	≤ 100 colonies	Passed
<i>Enterococcus faecalis</i> WDCM 0008	<10 colonies	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department


Huynh Bich Tram



Doan Duc Dai