

## CERTIFICATE OF ANALYSIS

### PRODUCT INFORMATION

Product: **BacterTube™ Sabourand Dextrose Broth**  
 Product Code: **08051**  
 Batch Number: **211224**  
 Expiration Date: **210625**  
 Storage Conditions: **2°C – 8°C**  
 Method: **Quality testing procedure: QT/KT/03.01**  
 Execution Date: **21/12/2024**

### RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform, clear, no sediment, no roughness, and no stratification	Passed
Color	Clear, amber solution	Passed
pH	5,6 ± 0,2	Passed
Tube Height (mm)	100 ± 2	Passed
Tube Diameter (mm)	16 ± 0,5	Passed
Volume (mL)	5 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
<b>Culture conditions: 72 hours, incubated at 30 – 35°C, inoculum ≤ 10<sup>2</sup> CFU/mL</b>		
<i>Candida albicans</i> ATCC 10231	Good Growth	Passed
<b>Culture conditions: 48 - 72 hours, incubated at 20 – 25°C, inoculum ≤ 10<sup>2</sup> CFU/mL</b>		
<i>Saccharomyces cerevisiae</i> ATCC 9763	Good Growth	Passed
<i>Aspergillus brasiliensis</i> ATCC 16404	Good Growth	Passed

### CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

**Quality Inspection Personnel**

**Quality Management Department**

  
**Huynh Bich Tram**

  
**Doan Duc Dai**

CTY TNHH TBKH LABONE  
 QC  
 TESTED