

BTL/BM/HD/QTTN/01.03

## CERTIFICATE OF ANALYSIS

### PRODUCT INFORMATION

Product: **BacterTube™ Cary Blair Broth**  
 Product Code: **08053**  
 Batch Number: **211224**  
 Expiration Date: **210625**  
 Storage Conditions: **2°C – 8°C**  
 Method: **Quality testing procedure: QT/KT/03.01**  
 Execution Date: **21/12/2024**

### RESULTS


CRITERIA	STANDARD	RESULTS
Appearance	Uniform, clear, no sediment, no roughness, and no stratification	Passed
Color	Amber, transparent	Passed
pH	8,4 ± 0,2	Passed
Tube Height (mm)	100 ± 2	Passed
Tube Diameter (mm)	16 ± 0,5	Passed
Volume (mL)	5 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
<b>Culture conditions: 18 – 24 hours, incubated at 35 – 37°C, inoculum ≤ 10<sup>2</sup> CFU/mL, aerobic</b>		
<i>Salmonella typhimurium</i> ATCC 14028	Good Growth	Passed
<i>V.parahaemolyticus</i> ATCC 17802	Good Growth	Passed
<i>Shigella flexneri</i> ATCC 12022	Good Growth	Passed

### CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

 (Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

**Quality Inspection Personnel**
**Quality Management Department**

**Huynh Bich Tram**

**Doan Duc Dai**