

## CERTIFICATE OF ANALYSIS

### PRODUCT INFORMATION

Product: **LabCell™ Dulbecco's Phosphate Buffered Saline – DPBS**  
Product Code: **13013**  
Batch Number: **071124**  
Expiration Date: **071126**  
Storage Conditions: **2°C – 8°C**  
Method: **Quality testing procedure: QT/KT/03.01**  
Execution Date: **07/11/2024**

### RESULTS

| CRITERIA    | STANDARD  | RESULTS |
|-------------|---|---------|
| Appearance  | Homogeneous solution, no residue, no stratification                             | Passed  |
| Color       | Colorless, transparent  | Passed  |
| pH          | $6,9 \pm 0,2$   | Passed  |
| Volume (mL) | $500 \pm 2$   | Passed  |
| Sterility   | No bacterial growth after incubation at $35 - 37^\circ\text{C}$ for 18 – 24 hrs | Passed  |

### CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, October 09, 2024

Quality Inspection Personnel

Trâm  
Huỳnh Bích Trâm

Quality Management Department



Đoàn Đức Đại