

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **LabCell™ GMEM**
Product Code: **13001**
Batch Number: **071124**
Expiration Date: **071126**
Storage Conditions: **2°C – 8°C**
Method: **Quality testing procedure: QT/KT/03.01**
Execution Date: **07/11/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Homogeneous solution, no residue, no stratification	Passed
Color	Bright red, transparent	Passed
pH	7,5 ± 0,3	Passed
Volume (mL)	500 ± 2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, October 09, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm



Đoàn Đức Đại