

## CERTIFICATE OF ANALYSIS

### PRODUCT INFORMATION

Product: **TraceMedia™ Chromagar/ BacterChrom™ UTI Agar**  
 Product Code: **TM10722.010/ 01012**  
 Batch Number: **181124**  
 Expiration Date: **180325**  
 Storage Conditions: **2°C – 8°C**  
 Method: **Quality testing procedure: QT/KT/03.01**  
 Execution Date: **18/11/2024**

### RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Smooth, uniform surface, no roughness, no layering	Passed
Color	Milky white agar, opaque	Passed
pH	6,8 ± 0,2	Passed
Agar Thickness (mm)	4 ± 1	Passed
Dish Diameter (mm)	90 ± 0,5	Passed
Volume (mL)	20 ± 2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
<b>Culture conditions: 18 – 24 hours, incubated at 35 – 37°C, inoculum ≤ 10<sup>3</sup> – 10<sup>4</sup> CFU/mL</b>		
<i>E. coli</i> ATCC 35218	Good growth, pink colonies	Passed
<i>S. aureus</i> ATCC 25923	Good growth, pale yellow colonies	Passed
<i>E. faecalis</i> ATCC 29212	Good growth, blue colonies	Passed

### CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, November 19, 2024

**Quality Inspection Personnel**

**Quality Management Department**



**Huynh Bich Tram**



**Doan Duc Dai**