

## CERTIFICATE OF ANALYSIS

### PRODUCT INFORMATION

Product: **BacterChrom™ Mycoplasma**  
 Product Code: **01034**  
 Batch Number: **211224**  
 Expiration Date: **210325**  
 Storage Conditions: **2°C – 8°C**  
 Method: **Quality testing procedure: QT/KT/03.01**  
 Execution Date: **21/12/2024**

### RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Smooth, uniform surface, no roughness, no layering	Passed
Color	Milky white agar, opaque	Passed
pH	7,6 ± 0,2	Passed
Agar Thickness (mm)	4 ± 1	Passed
Dish Diameter (mm)	90 ± 0,5	Passed
Volume (mL)	20 ± 2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 72 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
<b>Culture conditions: 3 – 7 days, incubated at 35 – 37°C, 5 – 10% CO<sub>2</sub>, inoculum ≤ 10<sup>3</sup> – 10<sup>4</sup> CFU/mL</b>		
<i>M. bovis</i> ATCC 25523	Good growth	Passed
<i>S. aureus</i> ATCC 25923	Inhibited, no growth	Passed

### CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 28, 2024

**Quality Inspection Personnel**

**Quality Management Department**



**Huynh Bich Tram**




**Doan Duc Dai**