

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **BacterBag™ Dry Half-Fraiser Broth**
 Product Code: **10014**
 Batch Number: **211224**
 Expiration Date: **211227**
 Storage Conditions: **8°C – 25°C**
 Method: **Quality testing procedure: QT/KT/03.01**
 Execution Date: **21/12/2024**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Fine powder, uniform in color	Passed
Color	Creamy white, homogeneous powder	Passed
pH	7.2 ± 0.2	Passed
Weight (Kg)	5	Passed
Volume (L)	5 ± 0.2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 18 – 24 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 18 – 24 hours, incubated at 35 – 37°C, inoculum ≤ 10³ – 10⁴ CFU/mL		
<i>L. monocytogenes</i> ATCC 13932	Good growth	Passed
<i>E. coli</i> ATCC 35218	Inhibited, no growth	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huynh Bich Tram



Doan Duc Dai