

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **BacterContact™ Pro Chocolate + LTH**
 Product Code: **12022**
 Batch Number: **211224**
 Expiration Date: **210925**
 Storage Conditions: **2°C – 8°C**
 Method: **Quality testing procedure: QT/KT/03.01**
 Execution Date: **21/12/2024**

PRODUCT COMPOSITION

SUBSTANCE	WEIGHT
Special Peptone	23,0 g
Starch	0,1 g
Sodium Chloride	5,0 g
Agar	15,0 g
Horse Blood	50 – 70 mL
Polysorbate (Tween) 80	5,0 mL
Lecithin	0,7 g
Histidine	0,5 g

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Smooth, uniform surface, no roughness, no layering	Passed
Specification (Cellulose film)	3 layers	Passed
Color	Brown agar, Cloudy	Passed
pH	7,3 ± 0,2	7,36
Agar Thickness (mm)	3 ± 1	3,1
Dish Diameter (mm)	60 ± 0,5	60,25
Volume (mL)	13,5 ± 0,2	13,5
Sterility	No bacterial growth after incubation at 35 – 37°C, for 24 hrs	Passed

PERFORMANCE	STANDARD (Pr)	RESULTS
Culture conditions: 24 hrs, incubated at 35 – 37°C, inoculum ≤ 10² CFU/mL		
<i>H. influenza</i> ATCC 33533	≥ 70 %	98 %
<i>N. meningitidis</i> ATCC 13090	≥ 70 %	102 %

BTL/BM/HD/QTTN/01.03


CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huynh Bich Tram



Doan Duc Dai