

BM/KT/03/02

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **SwabCollect™ Neutralizing Buffer, 5 mL**
Product Code: **4108007**
Batch Number: **120525**
Expiration Date: **121126**
Storage Conditions: **4°C – 25°C**
Method: **Quality testing procedure: QT/KT/03.01**
Execution Date: **12/05/2025**

QUALITY TEST RESULTS

CRITERIA	STANDARD	RESULTS
Appearance, color	Homogeneous solution, no phase separation, light yellow color, transparent	Passed
pH	7,2 ± 0,2	Passed
Volume (mL)	5 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 72 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 48 hours, incubated at 35–37°C, inoculum ≤ 10² CFU/mL		
<i>Staphylococcus aureus</i> ATCC 6538 (WDCM 00032)	Good growth	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the following link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, May 16, 2025

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm



Doãn Đức Đại