

BM/KT/03.02

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **BacterBag[™] Maximum Recovery Diluent**
 Product Code: **10004**
 Batch Number: **211224**
 Expiration Date: **211225**
 Storage Conditions: **4°C – 25°C**
 Method: **Quality testing procedure: QT/KT/03.01**
 Execution Date: **21/12/24**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform, clear, no sediment, no roughness, and no stratification	Passed
Color	Light yellow coloured clear solution without any precipitate	Passed
pH	7,0 ± 0,2	Passed
Weight (Kg)	5	Passed
Volume (L)	5 ± 0,5	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 72 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 18 – 24 hours, incubated at 35 – 37°C, inoculum ≤ 10² CFU/mL		
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	Good growth	Passed
<i>Staphylococcus aureus</i> ATCC 25923 (WDCM 00034)	Good growth	Passed


CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 25, 2024

Quality Inspection Personnel

Quality Management Department



Huynh Bich Tram



Doan Duc Dai