

BM/KT/03.02

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **SwabCollect™ Cary Blair with Charcoal, 5mL**
Product Code: **4106009**
Batch Number: **040825**
Expiration Date: **040827**
Storage Conditions: **4°C – 25°C**
Method: **Quality testing procedure: QT/KT/03.01**
Execution Date: **04/08/2025**

RESULTS

| CRITERIA | STANDARD | RESULTS |
|-------------|--|---------|
| Appearance | Uniform, no sediment, no roughness, and no stratification | Passed |
| Color | Black coloured opaque gel | Passed |
| pH | 8,4 ± 0,2 | Passed |
| Volume (mL) | 5 ± 0,2 | Passed |
| Sterility | No bacterial growth after incubation at 35 – 37°C for 72 hrs | Passed |

| PERFORMANCE | STANDARD | RESULTS |
|--|-------------|---------|
| Culture conditions: 18 – 24 hours, incubated at 35 – 37°C, inoculum ≤ 10² CFU/mL | | |
| <i>Escherichia coli</i> ATCC 25922 | Good growth | Passed |
| <i>Neisseria meningitidis</i> ATCC 13090 | Good growth | Passed |
| <i>Salmonella Typhimurium</i> ATCC 14028 | Good growth | Passed |
| <i>Shigella flexneri</i> ATCC 12022 | Good growth | Passed |

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, Aug 08, 2025

Quality Inspection Personnel

Quality Management Department



Tran Ngoc Sang


Doan Duc Dai