

BM/KT/03.02

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **SwabCollect™ Maximum Recovery Diluent, 1 mL**
 Product Code: **4108033**
 Batch Number: **120525**
 Expiration Date: **121126**
 Storage Conditions: **4°C – 25°C**
 Method: **Quality testing procedure: QT/KT/03.01**
 Execution Date: **12/05/2025**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform, clear, no sediment, no roughness, and no stratification	Passed
Color	Light yellow coloured clear solution without any precipitate	Passed
pH	7,0 ± 0,2	Passed
Volume (mL)	1 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 72 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 18 – 24 hours, incubated at 35 – 37°C, inoculum ≤ 10² CFU/mL		
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	Good growth	Passed
<i>Staphylococcus aureus</i> ATCC 25923 (WDCM 00034)	Good growth	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, May 16, 2025

Quality Inspection Personnel

Quality Management Department



Huynh Bich Tram




Doan Duc Dai