

BM/KT/03.02

## CERTIFICATE OF ANALYSIS

### PRODUCT INFORMATION

Product: **BacterContact™ Sabouraud Dextrose Agar + 0.1% Penase + Neutralizing**  
 Product Code: **4109025**  
 Batch Number: **040825**  
 Expiration Date: **040226**  
 Storage Conditions: **2°C – 8°C**  
 Method: **Quality testing procedure: QT/KT/03.01**  
 Execution Date: **04/08/2025**

### RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Smooth, uniform surface, no roughness, no layering	Passed
Color	Slightly opalescent, amber	Passed
pH	5,6 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 72 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
<b>Culture conditions: 72 hours, incubated at 20 – 25°C, inoculum ≤ 10<sup>2</sup> CFU/mL</b>		
<i>Candida albicans</i> ATCC 10231	Good growth (P <sub>R</sub> ≥ 0,7)	Passed
<i>Aspergillus brasiliensis</i> ATCC 16404	Good growth (P <sub>R</sub> ≥ 0,7)	Passed
<b>Culture conditions: 24 hours, incubated at 30 – 35°C, inoculum ≤ 10<sup>2</sup> CFU/mL</b>		
<i>Staphylococcus aureus</i> ATCC 25923	Good growth (P <sub>R</sub> ≥ 0,7)	Passed
<i>Escherichia coli</i> ATCC 25922	Good growth (P <sub>R</sub> ≥ 0,7)	Passed

### CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

 (Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/> )

Ho Chi Minh City, Aug 08, 2025

**Quality Inspection Personnel**
**Quality Management Department**



**Tran Ngoc Sang**
**Doan Duc Dai**