

BM/KT/03.02

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **BacterContact™ Tryptic Soy Agar + Neutralizing**
 Product Code: **4109034**
 Batch Number: **040825**
 Expiration Date: **040226**
 Storage Conditions: **2°C – 8°C**
 Method: **Quality testing procedure: QT/KT/03.01**
 Execution Date: **04/08/2025**

RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Smooth, uniform surface, no roughness, no layering	Passed
Color	Slightly opalescent, amber	Passed
pH	7,3 ± 0,2	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 72 hrs	Passed

PERFORMANCE	STANDARD	RESULTS
Culture conditions: 48 – 72 hours, incubated at 20 – 25°C, inoculum ≤ 10² CFU/mL		
<i>Candida albicans</i> ATCC 10231	Good growth, 50-200% recovery	Passed
<i>Aspergillus brasiliensis</i> ATCC 16404	Good growth, 50-200% recovery	Passed
Culture conditions: 18 – 24 hours, incubated at 30 – 35°C, inoculum ≤ 10² CFU/mL		
<i>Staphylococcus aureus</i> ATCC 25923	Good growth, 50-200% recovery	Passed
<i>Escherichia coli</i> ATCC 25922	Good growth, 50-200% recovery	Passed
<i>Pseudomonas aeruginosa</i> ATCC 9027	Good growth, 50-200% recovery	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, Aug 08, 2025

Quality Inspection Personnel

Quality Management Department



Tran Ngoc Sang




Doan Duc Dai