

BTL/BM/HD/QTTN/01.03

CERTIFICATE OF ANALYSIS

PRODUCT INFORMATION

Product: **TraceTest™ LABONE HP test/ BacterTest™ LABONE HP test**
Product Code: **TT0122.050/ CL0123**
Batch Number: **211224**
Expiration Date: **210625**
Storage Conditions: **2°C – 8°C**
Method: **Quality testing procedure: QT/KT/03.01**
Execution Date: **21/12/24**

QUALITY TEST RESULTS

CRITERIA	STANDARD	RESULTS
Appearance	Uniform, no residue, no impurities, and no stratification	Passed
Color	Yellow - orange	Passed
Volume (mL)	250 µL	Passed
Sterility	No bacterial growth after incubation at 35 – 37°C for 16 – 24 hrs	Passed
Time	20 minutes	Passed

CONCLUSION: THE PRODUCT MEETS THE QUALITY REQUIREMENTS

*Note: Sensitivity: 100%; Specificity: $\geq 99,5\%$

(Please download the certificate of analysis at the link: <https://labone.com.vn/certificate-of-analysis-coa/>)

Ho Chi Minh City, December 23, 2024

Quality Inspection Personnel

Quality Management Department



Huỳnh Bích Trâm



Doãn Đức Đại